



03:2019-04 W000004

READ CAREFULLY BEFORE USE

INDICATIONS FOR USE

Finishing natural teeth and restorations

PRECAUTIONS

- 1. This product is intended for used by dental professionals only.
- 2. Do not use this product for any purpose other than specifically outlined in these instructions for use.

DIRECTIONS FOR USE

- 1. Sterilize diamond burs before each use.
 - ➤ Sterilization method: autoclave this product for 3 min at 134 °C (273 °F).

Cycle Type	Holding time	Temperature	Dry time
Pre-vacuum (3 times)	3'00"	134°C/ 273°F	10'00"

- Remove any contaminants, using washer disinfector.
- Do not immerse diamond burs into chlorine disinfectants.
- ➤ Follow the instructions for use of washer disinfector and sterilizer, respectively.
- 2. Mount a diamond bur on a dental handpiece (dental air turbine, micro motor etc.).
 - ➤ Insert the shank completely into the chuck of the handpiece, following the manufacturer's instructions.
 - Perform test run to check safety and smooth operation without run-out prior to use.
 - Make sure that head of a diamond bur is in good condition before use. Do not use damaged or defective instruments i.e. corrosion, surface flaw, deformation, dirt, exposure of shank due to repeated use.
- 3. Use following the standard manner.
 - > Never exceed the maximum permissible rotation speed.
 - ➤ Use local dust extractor, protective mask, etc while finishing to avoid adverse effect of dust to the human body.
 - Use protective equipment such as glasses while using.
 - Use intermittently with gentle touch and water spray.
 - Avoid any use at acute angles or with excessive pressure to thin and long abrasives. They are fragile and may bend or fracture under excessive loads.
- 4. For reuse after using them intraorally, remove any contaminants immediately using washer disinfector and sterilize.

Manufacturer

SHOFU INC

11 Kamitakamatsu-cho, Fukuine, Higashiyama-ku, Kyoto 605-0983, Japan www.shofu.com

SHOFU DENTAL ASIA-PACIFIC PTE.LTD. 10 Science Park Road, #03-12, The Alpha, Science Park II, Singapore 117684